

QC Tabular and Graphic Report

Family Health of Marco
935 Sundrop Ct 1192
Marco Island, FL 34145

Abbott Emerald

Cell Dyn H Control


Lot # H4244

Exp: 12/31/2019 12:00:00 AM

RBC

Number of runs: 44

Expected	Actual
Mean 5.200	5.195
SD 0.3000	0.0964
CV 1.86	



Run Date	Value	Status	SD	Out	Tech	Comments
09/02/2019	5.230	A	0.0600		SB	
09/02/2019	5.170	A	0.0000		SB	
09/02/2019	5.120	A	-0.0500		SB	
09/03/2019	5.210	A	0.0400		SB	
09/03/2019	5.250	A	0.0800		SB	
09/03/2019	5.330	A	0.1600		SB	
09/04/2019	5.250	A	0.0800		SB	
09/04/2019	5.230	A	0.0600		SB	
09/04/2019	5.330	A	0.1600		SB	
09/05/2019	5.250	A	0.0800		SB	
09/05/2019	5.180	A	0.0100		SB	
09/05/2019	5.210	A	0.0400		SB	
09/08/2019	5.200	A	0.0300		SB	
09/08/2019	5.220	A	0.0500		SB	
09/08/2019	5.220	A	0.0500		SB	
09/09/2019	5.250	A	0.0800		SB	
09/09/2019	5.260	A	0.0900		SB	
09/09/2019	5.310	A	0.1400		SB	
09/10/2019	5.290	A	0.1200		SB	
09/10/2019	5.250	A	0.0800		SB	
09/10/2019	5.310	A	0.1400		SB	
09/11/2019	5.180	A	0.0100		SB	
09/11/2019	5.050	A	-0.1200		SB	
09/11/2019	5.180	A	0.0100		SB	
09/12/2019	5.170	A	0.0000		SB	
09/12/2019	5.280	A	0.1100		SB	
09/12/2019	5.210	A	0.0400		SB	
09/15/2019	5.090	A	-0.0800		SB	
09/15/2019	5.280	A	0.1100		SB	
09/15/2019	5.310	A	0.1400		SB	
09/16/2019	5.210	A	0.0400		SB	
09/16/2019	5.220	A	0.0500		SB	
09/16/2019	5.240	A	0.0700		SB	

Abbott Emerald

Cell Dyn H Control

Lot # H4244

Exp: 12/31/2019 12:00:00 AM

09/17/2019	5.200	A	0.0300	SB
09/17/2019	5.130	A	-0.0400	SB
09/17/2019	4.950	A	-0.2200	SB
09/19/2019	5.180	A	0.0100	SB
09/22/2019	5.260	A	0.0900	SB
09/23/2019	5.200	A	0.0300	SB
09/24/2019	5.200	A	0.0300	SB
09/25/2019	5.050	A	-0.1200	SB
09/26/2019	4.940	A	-0.2600	SB
09/29/2019	5.060	A	-0.1400	SB
09/30/2019	4.930	A	-0.2700	SB